



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Scalemen of Florida

3600 Hacienda Boulevard, Ste A, Davie, FL 33314

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Calibration of Weighing Devices ***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

September 07, 2006

Issue Date:

July 31, 2022

Expiration Date:

October 31, 2024

Accreditation No.:

59400

Certificate No.:

L22-519

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

Scalemen of Florida

3600 Hacienda Blvd., Ste A, Davie, FL 33314
 Contact Name: Kim Hammock Phone: 954-370-8800

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Lab Balances ^{FO}	Up to 5 kg	$(1.16 \times 10^{-3} + 1.16 \times 10^{-4} \text{Wt}) \text{ g}$	Class 1 Weights Euromet Calibration Guide Number 18 Version 4
Scales and Balances ^{FO}	Up to 150 lb	$(1.15 \times 10^{-2} + 2.80 \times 10^{-5} \text{Wt}) \text{ lb}$	Class F Weights NIST HB 44
	151 lb to 5 000 lb	$(0.12 + 5.1 \times 10^{-5} \text{Wt}) \text{ lb}$	
	1 001 lb to 120 000 lb	$(0.15 + 1.1 \times 10^{-4} \text{Wt}) \text{ lb}$	
Weights ^F	1 g to 5 g	0.000 29 g	Class 1 Weight OIML R111
	10 g	0.000 3 g	
	20 g to 30 g	0.000 41 g	
	50 g	0.000 54 g	
	100g	0.000 82 g	
	200g	0.001 9 g	
	300 g	0.59 g	
	500 g to 2 000 g	0.71 g	
	1 lb	0.002 7 lb	
	2 lb	0.002 2 lb	
	5 lb	0.003 9 lb	
	10 lb	0.006 8 lb	
	25 lb	0.013 lb	
	50 lb	0.014 lb	
	1 g to 5 g	0.000 042 g	
	10 g	0.000 043 g	
	20 g to 30 g	0.000 047 g	
	50 g	0.000 067 g	
	100 g	0.000 093 g	
	200 g	0.000 17 g	
	300 g	0.001 8 g	
	500 g to 1 000 g	0.002 0 g	
	2 000 g	0.017 g	
Weights Class 7 ^F	1 lb.	0.000 16	NIST 105-1 Class F OIML R111
	2 lbs.	0.000 1	
	5 lbs.	0.000 14	
	10 lbs.	0.000 14	
	25 lbs.	0.005 6	
	50 lbs.	0.005 6	



Certificate of Accreditation: Supplement

Scalemen of Florida

3600 Hacienda Blvd., Ste A, Davie, FL 33314
Contact Name: Kim Hammock Phone: 954-370-8800

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Truck & Rail Scale ^O	1 000 lbs to 120 000 lbs.	(23.0 626 + 3.18 x 10 ⁻⁵) wt.	Class F Weights NIST Handbook 44
Force Gauges ^{FO}	Up to 100 lbf. 100.1 lbf. to 500 lbf	0.027 lbf. 0.11 lbf.	Class F Weights Manufacturer Manual WI-120 Force Measurement
Crane Scales Dynamometers Load Cells (Tension & Compression) ^F	Up to 20 000 lbf. 20 001 lbf to 40 000 lbf. 40 001 lbf to 60 000 lbf. 60 001 lbf to 80 000 lbf. 80 001 lbf to 100 000 lbf.	5.4 lbf 5.8 lbf. 6.5 lbf. 7.4 lbf. 8.3 lbf.	Master Load Cell ASTM E-4:2021 ASTM E-74:20218 WI-120 Force Measurement

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.